



WOBURN ULTIMATE PROFILING BED



Product Code:

HLB798.12
HLB798.14

Description:

The Woburn Ultimate Profiling Bed is our higher capacity bed within the Woburn range. Featuring triple aluminium side rails and under-bed strengthening, the Ultimate excels in patient support and ease of infection control. Available in 1200mm wide, with a 60 Stone weight capacity, this bed is able to support most bariatric patients. Comfort adjustments are made with ease due to the powerful lifting components and the high quality electronic control system.

The latest version of the Woburn Ultimate profiling bed now utilises a lock built into the side rail fingers themselves to operate the side rails, as seen on our Woburn Community profiling bed.

This is a simple and robust way to drop side rails, and the mechanism can be replaced quickly and efficiently by replacing the finger cartridge if required. This increases bed lifespan and improves speed of access to the bed.

Safe Working Load:

380kg (60 Stone)

Max. Patient Weight:

300kg

Dimensions:

2195 x 1340mm (1200mm mattress base)

2195 x 1540mm (1400mm mattress base)

Max. Mattress Depth:

260mm

Adjustable Height:

275 - 690mm

Angle of Back Rest:

Up to 70 degrees

Angle of Knee Break:

Up to 30 degrees

Bed Weight:

157kg + 16kg with rails

Side Rails

Description:

Aluminium side rails with plastic end caps

Dimensions:

1973 x 70 x 28mm

Key Information:

4 Section Profiling
Lockable Handset
Electric Backrest and Knee Break with Mechanical
Lower Leg Support
Wraparound Wooden Head and Footboards
Aluminium Side Rails
Auto-Regression as standard
Trendelenburg Function as standard
200mm Extension - Optional
Low Voltage – 29v; 9v Emergency Battery Backup
In-built Stability System
Comprehensive Manual
CE marked
Compatible with environmental control system
Comes with a 5-year warranty

Supplier Contact Details:

t: +44(0)1709 377172 e: enquiries@harvesthealthcare.co.uk w: www.harvesthealthcare.co.uk
Sheaf house, Bradmarsh Way, Bradmarsh Business Park, Rotherham UK, S60

Important Safety Notice

<Ufj Ygh<YUthVUfY @tX gi ddYgYei Jda YbhUbX UddJubVWg hc Vwj YfUfub[YcZa YX]WU VebXJHcbg" GdYVWVWkC]W cZyei Jda Ybhig hY gcY fYgdcbgJ]JmcZhY WbJWU dfUWVjcbYf cfVUfYf
cj YfgYYJb[JbXJ Jxi U dUfYbhVUfY" Rig_ UggYgga YbtrgdYVWVWc i gYfa i ghVY VUfYfYX ci hdfJcf hc hY i gYcZU dfeXi VUg"
PfcXi VUg UbX XUfU VebJbi U mi bXYf[c XYJ Ycda YbhUbX a UmhYfYzRY XYJ JUV Zca hY XYfUJg [J Yb"

Version 2: Aug 2018