

GEL CUSHION



Product Code:

P311C40401HF
P311C45431HF

Description:

- The Gel Cushion is made of high resilience, impermeable, viscoelastic polyurethane gel.
- Designed to enhance existing support surfaces.
- Supplied with a navy blue washable, water resistant cover.

Dimensions:

P311C40401HF: 400mm x 400mm x 25mm
P311C45431HF: 450mm x 430mm x 25mm

Weight of Cushion:

P311C40401HF: 3.9 kg
P311C45431HF: 4.7 kg

**User Weight and
Pressure Relief Risk:**

Up to 95 kg - High Risk
Up to 120kg - Medium Risk

Cover:

Anti-allergenic, antibacterial, antimicrobial, antifungal, two way stretch, vapour permeable, washable, welded seams, two side zip. User instructions are printed on the cover.

Fire Retardancy:

NF EN ISO 597 - 1 & 2

Cleaning Guidance:

The cushion foam may be hand washed with mild detergent. Do not immerse in water. Allow to dry naturally. Cover should be wiped down with hot water and mild detergent or machine washed up to 95°C and tumble dried. For a comprehensive list of tested cleaning products see our website.

Warranty

In terms of Harvest Healthcare Guarantee and Warranties the Gel Cushion is guaranteed for three years from date of purchase.

Exclusions:

1. Removable cushion covers are guaranteed for one year
2. Damage to fabric covers through incorrect use and penetration by sharp instruments is excluded

Supplier Contact Details:

t: +44(0)1709 377172 e: enquiries@harvesthealthcare.co.uk w: www.harvesthealthcare.co.uk
Sheaf house, Bradmarsh Way, Bradmarsh Business Park, Rotherham UK, S60

Important Safety Notice

Harvest Healthcare Ltd supplies equipment and appliances to cover a range of medical conditions. Specific choice of equipment is the sole responsibility of the clinical practitioner or carer overseeing individual patient care. Harvest Healthcare Ltd supplies a range of products such as foam mattresses, active air mattresses, fall protection mats and bed rail protection bumpers. These products are usually covered in vapour permeable PU fabric that could cause risk of suffocation in the event of prolonged facial contact.
(Mattress risk categories are to be used as a guide only)

Risk assessments specific to user must be carried out prior to the use of all products. Products and data continually undergo development and may therefore deviate from the details given.

Version 2: Nov. 2017