

ELEXO HYGIENE CHAIR



Description:

- The Elexo provides comfort and security for both carers and users. Its variable height positioning helps to minimise carer's back strain.
- The electric height-adjustability ensures that the seat reaches a maximum tilting angle of 17° when at its maximum height, thereby creating the optimum seating balance for the user.

Safe Working Load: 180kg (28 Stone)

Adjustable Height: 50 - 105mm

Seat Width: 48cm

Seat Depth: 46cm

Wedge Setting: 0° - 17°

Length Basic Frame: 83cm

Width Basic Frame: 68cm

Weight: 50kg

Electrical Power: 24V

Battery Capacity: 7Ah

Charger: 220V 50/60 Hz 500mA

Key Features:

Simple electric high/low adjustment.
Hand control with battery indicator.
Sleep mode setting after 60 minutes.
Maximum weight 180kg.
Lowest position 50cm, highest position 105cm.
Frame width 68cm, depth 83cm.
Tilt adjustment to a maximum of 17 degrees.
Upholstered seat with toilet opening.
Inlay for the toilet opening.
Foldable upholstered armrests with front lock to enable sideways transfer.
Ergonomically formed seat and back rests.
Fold-out footrests with calf protection.
Fold-away and removable leg supports.
Rail for toilet bucket or bedpan.
Durable castors with swivel & lock features.
Extremely stable due to the fitting of 2 lifting motors.
Simple to recharge, with charger included.



Ergonomically formed seat with inlay



Ergonomically formed seat for toilet use



Plastic bedpan



Independent, foldup upholstered armrests with front lock



Handcontrol with battery indicator



Charge point with magnetic coupling

Supplier Contact Details:

t: +44(0)1709 377172 e: sales@harvesthealthcare.co.uk; w: www.harvesthealthcare.co.uk Sheaf house, Bradmarsh Way, Bradmarsh Business Park, Rotherham UK, S60 1BW

Important Safety Notice

Harvest Healthcare Ltd supplies equipment and appliances to cover a range of medical conditions. Specific choice of equipment is the sole responsibility of the clinical practitioner or carer overseeing individual patient care. Risk assessments specific to user must be carried out prior to the use of all products.

Products and data continually undergo development and may therefore deviate from the details given.

Version 1: Mar 2019