

VISCOFLEX MULTI-BEARING MATTRESS



Product Codes: P161MFPPMHF

Description:

- Very high density viscoelastic foam
- Added multibearing surface
- Uniquely shaped to accommodate bed accessories
- Four-way stretch, waterproof, anti-bacterial PU coated cover with welded seams and a crib 5 fire rating
- Comes with a 3-year warranty
- High to very high risk

Dimensions: 1990 x 860 x 140mm

Max. User Weight: 20 Stone (130 kg)

Risk Level: High - Very High Risk

Cover: (Available Separately) Welded four way stretch vapour permeable, PU cover is water resistant, washable, hypoallergenic and has antibacterial properties.

Foam Density
80kg/m³ foam (visco upper layer)
40kg/m³ foam (firm base layer)

Fire Retardancy: BS 7177: 2008 crib 5

Cleaning Guidance

The cushion foam may be hand washed with mild detergent. Do not immerse in water. Allow to dry naturally. Cover should be wiped down with hot water and mild detergent or machine washed up to 95°C and tumble dried. For a comprehensive list of tested cleaning products see our website.

Warranty:

Subject to the terms of Harvest Healthcare Guarantee and Warranties the Viscoflex Evolutive Mattress is guaranteed for three years from date of purchase.

Exclusions:

1. Removable covers are guaranteed for one year
2. Damage to fabric covers through incorrect use and penetration by sharp instruments is excluded

Supplier Contact Details:

t: +44(0)1709 377172 e: enquiries@harvesthealthcare.co.uk; w: www.harvesthealthcare.co.uk
Sheaf house, Bradmarsh Way, Bradmarsh Business Park, Rotherham UK, S60 1BW

Important Safety Notice

Harvest Healthcare Ltd supplies equipment and appliances to cover a range of medical conditions. Specific choice of equipment is the sole responsibility of the clinical practitioner or carer overseeing individual patient care. Harvest Healthcare Ltd supplies a range of products such as foam mattresses, active air mattresses, fall protection mats and bed rail protection bumpers. These products are usually covered in vapour permeable PU fabric that could cause risk of suffocation in the event of prolonged facial contact. (Mattress risk categories are to be used as a guide only) Risk assessments specific to user must be carried out prior to the use of all products.

Products and data continually undergo development and may therefore deviate from the details given.

Version 4: Aug 2021