

DUKE ACTIVE MATTRESS



Product Code:

HR440

Description:

- The Duke Replacement Mattress is a state of the art, high quality system for the prevention and treatment of pressure ulcers in very high risk patients
- It has 30 cells with an integral, high density, foam base
- The slim 140mm profile ensures side rail compliance
- 4-way stretch welded PU cover which is vapour permeable, water resistant and has a Crib 5 fire rating

Mattress Dimensions:

1980 x 900 x 140mm

Weight:

9kg

Max. Patient Weight:

Harvest 250 Pump: 26 Stone (165kg)
Harvest 3 Pump: 39 Stone (247kg)

Pressure Relief Risk:

Very High

Pump Technical Specification:

Please see separate pump specification sheet

Cleaning Guidance:

Cover should be wiped down with hot water and mild detergent. If heavily soiled, the cover may be machine washed up to 85°C and tumble dried. For disinfection use sodium hypochlorite solution max 1000ppm. Use no bleach, biological or phenolic-based cleaners. Dry completely before storing.

Warranty

In terms of Harvest Healthcare Guarantee and Warranties the Harvest Duke mattress is guaranteed for three years from date of purchase.

Exclusions:

1. Removable covers are guaranteed for one year
2. Damage to fabric covers through incorrect use and penetration by sharp instruments is excluded
3. Components of pumps which are recommended for annual service, are guaranteed for one year.

Supplier Contact Details:

t: +44(0)1709 377172 e: enquiries@harvesthealthcare.co.uk w: www.harvesthealthcare.co.uk
Sheaf house, Bradmarsh Way, Bradmarsh Business Park, Rotherham UK, S60

Important Safety Notice

Harvest Healthcare Ltd supplies equipment and appliances to cover a range of medical conditions. Specific choice of equipment is the sole responsibility of the clinical practitioner or carer overseeing individual patient care. Harvest Healthcare Ltd supplies a range of products such as foam mattresses, active air mattresses, fall protection mats and bed rail protection bumpers. These products are usually covered in vapour permeable PU fabric that could cause risk of suffocation in the event of prolonged facial contact.

(Mattress risk categories are to be used as a guide only)

Risk assessments specific to user must be carried out prior to the use of all products. Products and data continually undergo development and may therefore deviate from the details given.

Version 2: Jun 2018