

# Active Seat Cushion



**Product Code:**

**HH20**

**Description:**

- 6 alternating cells
- Zipped, water resistant, four-way stretch cover
- Non-slip base
- Can be paired with our Hampton mattress for 24 hour pressure area care
- Compatible with our Harvest 1, Harvest 250 Pumps
- Very High Risk

**Dimensions:**

430(L) x 430(W) x 90(H)mm

**Weight of Cushion:**

1.4 kg

**User Weight:**

158kg/25st with Harvest 1 Pump - 160kg/25st with Harvest 250 Pump

**Pressure Relief Risk:**

Very High

**Number of Cells:**

6 Alternating

**Pump:**

HPU1

**Pump Weight:**

1.7 KG

**Operating Cycle Time:**

10 Minutes

**Voltage:**

230V 50Hz

**Fuse:**

T1A 250v

**Safety Standards:**

EN 60601-1: 2006 and A11:2011

**Warranty**

In terms of Harvest Healthcare Guarantee and Warranties the Active Seat Systems are guaranteed for three years from date of purchase.

**Exclusions:**

1. Removable cushion covers are guaranteed for one year
2. Damage to fabric covers through incorrect use and penetration by sharp instruments is excluded
3. Components of pumps which are recommended for annual service, are guaranteed for one year.

**Supplier Contact Details:**

t: +44(0)1709 377172

e: enquiries@harvesthealthcare.co.uk

w: www.harvesthealthcare.co.uk

Sheaf house, Bradmarsh Way, Bradmarsh Business Park, Rotherham UK, S60

**Important Safety Notice**

Harvest Healthcare Ltd supplies equipment and appliances to cover a range of medical conditions. Specific choice of equipment is the sole responsibility of the clinical practitioner or carer overseeing individual patient care. Harvest Healthcare Ltd supplies a range of products such as foam mattresses, active air mattresses, fall protection mats and bed rail protection bumpers. These products are usually covered in vapour permeable PU fabric that could cause risk of suffocation in the event of prolonged facial contact.

(Mattress risk categories are to be used as a guide only)

Risk assessments specific to user must be carried out prior to the use of all products. Products and data continually undergo development and may therefore deviate from the details given.

Version 4: July 2021